

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016239.D
 Acq On : 14 May 2020 18:17
 Operator : JC/SP
 Sample : L2643-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z4-DUP-0557-200513

Quant Time: May 15 04:13:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	240045	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	395573	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	397524	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	206101	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	146797	46.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.94%	
35) Dibromofluoromethane	5.47	113	118742	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
50) Toluene-d8	8.70	98	493834	51.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.28%	
62) 4-Bromofluorobenzene	11.12	95	200100	54.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.96%	
Target Compounds						
16) Acetone	2.42	43	2872	2.038	ug/l	96
27) cis-1,2-Dichloroethene	4.57	96	19703	6.481	ug/l	86
44) Trichloroethene	7.19	130	1415	0.350	ug/l	97
64) Tetrachloroethene	9.32	164	981	0.201	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016239.D
Acq On : 14 May 2020 18:17
Operator : JC/SP
Sample : L2643-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z4-DUP-0557-200513

Quant Time: May 15 04:13:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

