

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX051420\
 Data File : VX016245.D
 Acq On : 14 May 2020 20:33
 Operator : JC/SP
 Sample : L2644-06 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DRUM-COMP

Manual Integrations
 APPROVED

MMDadoda
 5/15/2020 12:29:28 PM

Quant Time: May 15 09:26:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	235815	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	394676	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	384295	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	195485	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	144338	46.51	ug/l	0.00
Spiked Amount			Recovery	=	93.02%	
35) Dibromofluoromethane	5.47	113	115631	46.24	ug/l	0.00
Spiked Amount			Recovery	=	92.48%	
50) Toluene-d8	8.70	98	486083	50.45	ug/l	0.00
Spiked Amount			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	11.12	95	196340	53.58	ug/l	0.00
Spiked Amount			Recovery	=	107.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	4386	3.168	ug/l	89
17) Carbon Disulfide	2.55	76	6036	0.616	ug/l	98
20) Methylene Chloride	2.84	84	2136	0.748	ug/l	90
25) 2-Butanone	4.65	43	3410	1.528	ug/l	100
36) 1,1-Dichloropropene	5.76	75	1281	0.330	ug/l #	83
68) m/p-Xylenes	10.34	106	1622	0.302	ug/l	83
78) n-propylbenzene	11.35	91	3890	0.235	ug/l	99
82) 4-Chlorotoluene	11.49	91	3106	0.269	ug/l	99
86) p-Isopropyltoluene	12.05	119	7836	0.604	ug/l	92
87) 1,3-Dichlorobenzene	12.01	146	2015	0.306	ug/l	89
88) 1,4-Dichlorobenzene	12.08	146	2620m	0.392	ug/l	
89) n-Butylbenzene	12.37	91	5634	0.476	ug/l	99
93) 1,2,4-Trichlorobenzene	13.63	180	3130	0.666	ug/l	95
94) Hexachlorobutadiene	13.77	225	2568	1.780	ug/l	94
95) Naphthalene	13.82	128	6826	0.449	ug/l	100
96) 1,2,3-Trichlorobenzene	14.01	180	2967	0.646	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016245.D
Acq On : 14 May 2020 20:33
Operator : JC/SP
Sample : L2644-06 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DRUM-COMP

Manual Integrations
APPROVED

MMDadoda
5/15/2020 12:29:28 PM

Quant Time: May 15 09:26:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

