

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016252.D
 Acq On : 15 May 2020 00:19
 Operator : JC/SP
 Sample : PB128866TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB128866TB

Quant Time: May 15 05:10:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	248190	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	419374	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	405790	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	204402	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	150899	46.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.40%	
35) Dibromofluoromethane	5.47	113	120607	45.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.78%	
50) Toluene-d8	8.70	98	514226	50.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.46%	
62) 4-Bromofluorobenzene	11.12	95	204611	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
Target Compounds						
16) Acetone	2.42	43	13459	9.236	ug/l	95
20) Methylene Chloride	2.84	84	5585	1.859	ug/l	91
43) Isopropyl Acetate	6.42	43	78590	10.295	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016252.D
Acq On : 15 May 2020 00:19
Operator : JC/SP
Sample : PB128866TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB128866TB

Quant Time: May 15 05:10:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

