

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016258.D
 Acq On : 15 May 2020 02:35
 Operator : JC/SP
 Sample : PB128866ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB128866ZHE#06

Quant Time: May 15 05:53:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	250385	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	422013	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	415684	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	213998	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	153444	46.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.14%	
35) Dibromofluoromethane	5.47	113	122735	45.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.82%	
50) Toluene-d8	8.70	98	523590	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
62) 4-Bromofluorobenzene	11.12	95	208869	53.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.62%	
Target Compounds						
16) Acetone	2.42	43	14205	9.663	ug/l	93
20) Methylene Chloride	2.84	84	16331	5.389	ug/l	96
43) Isopropyl Acetate	6.42	43	74003	9.633	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016258.D
Acq On : 15 May 2020 02:35
Operator : JC/SP
Sample : PB128866ZHE#06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB128866ZHE#06

Quant Time: May 15 05:53:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

