

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016261.D
 Acq On : 15 May 2020 03:43
 Operator : JC/SP
 Sample : L2640-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 WASTE-WATER-CAKE

Quant Time: May 15 06:25:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	246995	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	422814	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	416482	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	220643	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	154333	47.48	ug/l	0.00
Spiked Amount				50.000		
Recovery					=	94.96%
35) Dibromofluoromethane	5.47	113	123833	46.23	ug/l	0.00
Spiked Amount				50.000		
Recovery					=	92.46%
50) Toluene-d8	8.70	98	527641	51.12	ug/l	0.00
Spiked Amount				50.000		
Recovery					=	102.24%
62) 4-Bromofluorobenzene	11.12	95	212901	54.23	ug/l	0.00
Spiked Amount				50.000		
Recovery					=	108.46%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	81296	56.058	ug/l	96
18) Methyl Acetate	2.75	43	3820	1.071	ug/l	99
20) Methylene Chloride	2.84	84	6063	2.028	ug/l	94
43) Isopropyl Acetate	6.42	43	61342	7.970	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016261.D
 Acq On : 15 May 2020 03:43
 Operator : JC/SP
 Sample : L2640-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 WASTE-WATER-CAKE

Quant Time: May 15 06:25:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

