

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016265.D
 Acq On : 15 May 2020 05:14
 Operator : JC/SP
 Sample : PB128867TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB128867TB

Quant Time: May 15 07:12:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	258773	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	430942	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	426184	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	220713	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	155750	45.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.48%	
35) Dibromofluoromethane	5.47	113	126561	46.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.72%	
50) Toluene-d8	8.70	98	535697	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
62) 4-Bromofluorobenzene	11.12	95	216133	54.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.04%	
Target Compounds						
16) Acetone	2.42	43	15254	10.040	ug/l	95
18) Methyl Acetate	2.75	43	4014	1.074	ug/l	93
20) Methylene Chloride	2.84	84	5682	1.814	ug/l	95
43) Isopropyl Acetate	6.41	43	82839	10.560	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016265.D
Acq On : 15 May 2020 05:14
Operator : JC/SP
Sample : PB128867TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB128867TB

Quant Time: May 15 07:12:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

