

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016266.D
 Acq On : 15 May 2020 05:37
 Operator : JC/SP
 Sample : L2641-05 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OIL-WATER-SEPERATOR-2

Quant Time: May 15 07:15:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	249409	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	417590	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	418404	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	214373	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.04	65	152788	46.55	ug/l	0.00
Spiked Amount			Recovery	=	93.10%	
35) Dibromofluoromethane	5.47	113	121701	46.00	ug/l	0.00
Spiked Amount			Recovery	=	92.00%	
50) Toluene-d8	8.70	98	520603	51.07	ug/l	0.00
Spiked Amount			Recovery	=	102.14%	
62) 4-Bromofluorobenzene	11.12	95	213550	55.08	ug/l	0.00
Spiked Amount			Recovery	=	110.16%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016266.D
Acq On : 15 May 2020 05:37
Operator : JC/SP
Sample : L2641-05 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OIL-WATER-SEPERATOR-2

Quant Time: May 15 07:15:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

