

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051420\
 Data File : VX016256.D
 Acq On : 15 May 2020 01:50
 Operator : JC/SP
 Sample : PB128866ZHE#04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB128866ZHE#04

Quant Time: May 15 05:43:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	247505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	410261	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	408371	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	211696	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	151023	46.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.74%	
35) Dibromofluoromethane	5.47	113	120806	46.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.96%	
50) Toluene-d8	8.70	98	508998	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
62) 4-Bromofluorobenzene	11.12	95	204793	53.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.52%	
Target Compounds						
16) Acetone	2.42	43	14181	9.758	ug/l	91
20) Methylene Chloride	2.84	84	10818	3.611	ug/l	95
43) Isopropyl Acetate	6.42	43	74762	10.011	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051420\
Data File : VX016256.D
Acq On : 15 May 2020 01:50
Operator : JC/SP
Sample : PB128866ZHE#04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB128866ZHE#04

Quant Time: May 15 05:43:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.64 min Scan# 746

Delta R.T. 0.00 min

Lab File: VX016256.D

Acq: 15 May 2020 01:50

Instrument :

MSVOA_X

ClientSampleId :


PB128866ZHE#04

Tot Ion:168 Resp: 247505

Ion Ratio Lower Upper




168 100

99 56.3 45.0 67.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#16

Acetone

Concen: 9.758 ug/l

RT: 2.42 min Scan# 218

Delta R.T. -0.01 min

Lab File: VX016256.D


Acq: 15 May 2020 01:50

Tgt Ion: 43 Resp: 14181

Ion Ratio Lower Upper



43 100


58 36.4 25.4 38.0



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#20
Methylene Chloride
Concen: 3.611 ug/l
RT: 2.84 min Scan# 287
Delta R.T. 0.00 min
Lab File: VX016256.D
Acq: 15 May 2020 01:50




Instrument :
MSVOA_X
ClientSampleId :
PB128866ZHE#04


Tot Ion: 84 Resp: 10818
Ion Ratio Lower Upper
84 100
49 117.1 97.3 145.9
51 37.3 29.3 43.9
86 71.4 51.3 76.9



#33
1,2-Dichloroethane-d4
Concen: 46.370 ug/l
RT: 6.03 min Scan# 811
Delta R.T. 0.00 min
Lab File: VX016256.D
Acq: 15 May 2020 01:50

Tgt Ion: 65 Resp: 151023
Ion Ratio Lower Upper
65 100
67 53.5 0.0 104.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.84 min Scan# 943

Delta R.T. 0.01 min

Lab File: VX016256.D

Acq: 15 May 2020 01:50

Instrument :

MSVOA_X

ClientSampleId :

PB128866ZHE#04


Tot Ion:114 Resp: 410261

Ion Ratio Lower Upper

114 100

63 20.0 0.0 42.8




88 16.3 0.0 31.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 46.479 ug/l

RT: 5.47 min Scan# 719

Delta R.T. 0.00 min

Lab File: VX016256.D

Acq: 15 May 2020 01:50


Tgt Ion:113 Resp: 120806

Ion Ratio Lower Upper

113 100

111 104.4 81.6 122.4



192 21.5 17.0 25.4




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 10.011 ug/l

RT: 6.42 min Scan# 874

Delta R.T. 0.00 min

Lab File: VX016256.D

Acq: 15 May 2020 01:50

Instrument :

MSVOA_X

ClientSampleId :

PB128866ZHE#04


Tot Ion: 43 Resp: 74762

Ion Ratio Lower Upper

43 100

61 21.9 17.0 25.4




87 14.4 11.0 16.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0



#50

Toluene-d8

Concen: 50.825 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. 0.00 min

Lab File: VX016256.D


Acq: 15 May 2020 01:50

Tgt Ion: 98 Resp: 508998

Ion Ratio Lower Upper



98 100


100 67.2 53.7 80.5



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0





#62

4-Bromofluorobenzene

Concen: 53.763 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. 0.00 min

Lab File: VX016256.D

Acq: 15 May 2020 01:50

Instrument :

MSVOA_X

ClientSampleId :

PB128866ZHE#04


Tot Ion: 95 Resp: 204793

Ion Ratio Lower Upper

95 100

174 79.1 0.0 159.4

176 77.7 0.0 157.6




Abundance Ion 94.90 (94.60 to 95.60): VX016032.D (-1639) (-)

Ion 173.80 (173.50 to 174.50): VX016032.D (-1639) (-)

Ion 175.80 (175.50 to 176.50): VX016032.D (-1639) (-)

Time-->




Abundance Ion 94.90 (94.60 to 95.60): VX016256.D (-1597) (-)

Ion 173.80 (173.50 to 174.50): VX016256.D (-1597) (-)

Ion 175.80 (175.50 to 176.50): VX016256.D (-1597) (-)

Time-->



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.10 min Scan# 1478

Delta R.T. 0.00 min

Lab File: VX016256.D

Acq: 15 May 2020 01:50


Tgt Ion: 117 Resp: 408371

Ion Ratio Lower Upper

117 100

82 55.5 44.4 66.6

119 31.6 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): VX016032.D (-1470) (-)

Ion 82.00 (81.70 to 82.70): VX016032.D (-1470) (-)

Ion 118.90 (118.60 to 119.60): VX016032.D (-1470) (-)

Time-->




Abundance Ion 116.90 (116.60 to 117.60): VX016256.D (-1429) (-)

Ion 82.00 (81.70 to 82.70): VX016256.D (-1429) (-)

Ion 118.90 (118.60 to 119.60): VX016256.D (-1429) (-)

Time-->



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.06 min Scan# 1800

Delta R.T. 0.00 min

Lab File: VX016256.D

Acq: 15 May 2020 01:50

Instrument :

MSVOA_X

ClientSampleId :

PB128866ZHE#04


Tot Ion:152 Resp: 211696

Ion Ratio Lower Upper

152 100

115 57.5 41.3 124.1

150 155.4 0.0 352.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

