

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051421\
 Data File : VX021994.D
 Acq On : 14 May 2021 14:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 15 01:05:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.658	0.659	-0.2	97	0.00
3 P	Chloromethane	0.791	0.719	9.1	90	0.00
4 C	Vinyl Chloride	0.913	0.778	14.8#	85	0.00
5 T	Bromomethane	0.333	0.311	6.6	98	0.00
6 T	Chloroethane	0.475	0.504	-6.1	99	0.00
7 T	Trichlorofluoromethane	1.258	1.097	12.8	86	0.00
8 T	Diethyl Ether	0.528	0.467	11.6	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.640	0.643	-0.5	103	0.00
10 T	Methyl Iodide	0.760	0.613	19.3	74	0.00
11 T	Tert butyl alcohol	0.220	0.175	20.5	82	-0.01
12 CM	1,1-Dichloroethene	0.678	0.619	8.7#	94	0.00
13 T	Acrolein	0.110	0.074	32.7#	85	0.00
14 T	Allyl chloride	1.242	1.116	10.1	91	0.00
15 T	Acrylonitrile	0.413	0.357	13.6	86	-0.01
16 T	Acetone	0.369	0.312	15.4	86	-0.01
17 T	Carbon Disulfide	1.886	1.695	10.1	90	0.00
18 T	Methyl Acetate	1.023	0.914	10.7	90	-0.01
19 T	Methyl tert-butyl Ether	2.174	2.094	3.7	95	0.00
20 T	Methylene Chloride	0.875	0.736	15.9	97	0.00
21 T	trans-1,2-Dichloroethene	0.723	0.683	5.5	93	0.00
22 T	Diisopropyl ether	2.313	2.121	8.3	91	0.00
23 T	Vinyl Acetate	2.276	2.066	9.2	88	0.00
24 P	1,1-Dichloroethane	1.330	1.278	3.9	96	0.00
25 T	2-Butanone	0.635	0.545	14.2	84	0.00
26 T	2,2-Dichloropropane	1.190	1.108	6.9	94	0.00
27 T	cis-1,2-Dichloroethene	0.854	0.822	3.7	96	0.00
28 T	Bromochloromethane	0.608	0.595	2.1	96	0.00
29 T	Tetrahydrofuran	0.377	0.319	15.4	82	0.00
30 C	Chloroform	1.356	1.289	4.9#	95	0.00
31 T	Cyclohexane	1.087	1.054	3.0	97	0.00
32 T	1,1,1-Trichloroethane	1.190	1.152	3.2	98	0.00
33 S	1,2-Dichloroethane-d4	0.823	0.790	4.0	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.343	0.327	4.7	95	0.00
36 T	1,1-Dichloropropene	0.530	0.521	1.7	99	0.00
37 T	Ethyl Acetate	0.640	0.567	11.4	86	0.00
38 T	Carbon Tetrachloride	0.530	0.517	2.5	95	0.00
39 T	Methylcyclohexane	0.595	0.586	1.5	97	0.00
40 TM	Benzene	1.596	1.558	2.4	95	0.00
41 T	Methacrylonitrile	0.346	0.313	9.5	90	-0.01
42 TM	1,2-Dichloroethane	0.547	0.546	0.2	97	0.00
43 T	Isopropyl Acetate	1.083	0.965	10.9	88	0.00
44 TM	Trichloroethene	0.388	0.393	-1.3	98	0.00
45 C	1,2-Dichloropropane	0.418	0.401	4.1#	94	0.00
46 T	Dibromomethane	0.283	0.273	3.5	97	0.00
47 T	Bromodichloromethane	0.570	0.550	3.5	94	0.00
48 T	Methyl methacrylate	0.498	0.452	9.2	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051421\
 Data File : VX021994.D
 Acq On : 14 May 2021 14:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 15 01:05:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.009	10.0	84	0.00
50 S	Toluene-d8	1.308	1.280	2.1	96	0.00
51 T	4-Methyl-2-Pentanone	0.674	0.587	12.9	84	0.00
52 CM	Toluene	1.019	0.994	2.5#	95	0.00
53 T	t-1,3-Dichloropropene	0.665	0.636	4.4	93	0.00
54 T	cis-1,3-Dichloropropene	0.696	0.668	4.0	93	0.00
55 T	1,1,2-Trichloroethane	0.407	0.396	2.7	95	0.00
56 T	Ethyl methacrylate	0.715	0.659	7.8	92	0.00
57 T	1,3-Dichloropropane	0.697	0.685	1.7	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.350	0.332	5.1	90	0.00
59 T	2-Hexanone	0.519	0.442	14.8	81	0.00
60 T	Dibromochloromethane	0.439	0.407	7.3	91	0.00
61 T	1,2-Dibromoethane	0.425	0.409	3.8	93	0.00
62 S	4-Bromofluorobenzene	0.516	0.504	2.3	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.322	0.364	-13.0	111	0.00
65 PM	Chlorobenzene	1.159	1.159	0.0	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.426	0.416	2.3	95	0.00
67 C	Ethyl Benzene	2.136	2.138	-0.1#	95	0.00
68 T	m/p-Xylenes	0.816	0.813	0.4	93	0.00
69 T	o-Xylene	0.815	0.789	3.2	95	0.00
70 T	Styrene	1.355	1.344	0.8	91	0.00
71 P	Bromoform	0.324	0.297	8.3	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	4.521	4.475	1.0	95	0.00
74 T	N-amyl acetate	2.423	2.116	12.7	84	0.00
75 P	1,1,2,2-Tetrachloroethane	1.677	1.483	11.6	87	0.00
76 T	1,2,3-Trichloropropane	1.422	1.297	8.8	85	0.00
77 T	Bromobenzene	0.974	0.953	2.2	93	0.00
78 T	n-propylbenzene	5.356	5.277	1.5	93	0.00
79 T	2-Chlorotoluene	3.245	3.077	5.2	91	0.00
80 T	1,3,5-Trimethylbenzene	3.885	3.804	2.1	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.613	0.529	13.7	81	0.00
82 T	4-Chlorotoluene	3.756	3.644	3.0	91	0.00
83 T	tert-Butylbenzene	3.766	3.604	4.3	91	0.00
84 T	1,2,4-Trimethylbenzene	3.897	3.819	2.0	92	0.00
85 T	sec-Butylbenzene	4.624	4.628	-0.1	93	0.00
86 T	p-Isopropyltoluene	3.941	3.890	1.3	92	0.00
87 T	1,3-Dichlorobenzene	1.840	1.862	-1.2	96	0.00
88 T	1,4-Dichlorobenzene	1.903	1.860	2.3	93	0.00
89 T	n-Butylbenzene	3.634	3.670	-1.0	91	0.00
90 T	Hexachloroethane	0.712	0.663	6.9	89	0.00
91 T	1,2-Dichlorobenzene	1.816	1.796	1.1	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.404	0.326	19.3	85	0.00
93 T	1,2,4-Trichlorobenzene	1.094	1.102	-0.7	97	0.00
94 T	Hexachlorobutadiene	0.427	0.438	-2.6	97	0.00
95 T	Naphthalene	4.482	4.425	1.3	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051421\
Data File : VX021994.D
Acq On : 14 May 2021 14:46
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 15 01:05:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.109	1.098	1.0	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6