

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051424\
 Data File : VX041512.D
 Acq On : 14 May 2024 12:07
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: May 15 00:57:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	187913	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	276186	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	262890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135324	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	105472	51.558	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.120%		
35) Dibromofluoromethane	5.385	113	86787	46.370	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 92.740%		
50) Toluene-d8	8.647	98	318345	53.739	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 107.480%		
62) 4-Bromofluorobenzene	11.079	95	120764	52.662	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 105.320%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051424\
Data File : VX041512.D
Acq On : 14 May 2024 12:07
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: May 15 00:57:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

