

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051424\
 Data File : VX041514.D
 Acq On : 14 May 2024 12:53
 Operator : JC/MD
 Sample : PB160869TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB160869TB

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/15/2024
 Supervised By : Mahesh Dadoda 05/15/2024

Quant Time: May 15 00:58:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	168005	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	248790	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242027	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119435	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93037	50.869	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.740%			
35) Dibromofluoromethane	5.391	113	76181	45.185	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 90.380%			
50) Toluene-d8	8.647	98	293117	54.929	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 109.860%			
62) 4-Bromofluorobenzene	11.079	95	110212	53.353	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 106.700%			
Target Compounds						
16) Acetone	2.392	43	24832	27.348	ug/l	Qvalue 100
18) Methyl Acetate	2.709	43	2280	1.218	ug/l	93
20) Methylene Chloride	2.788	84	7692	4.844	ug/l	86
37) Ethyl Acetate	4.751	43	3639m	1.447	ug/l	
43) Isopropyl Acetate	6.342	43	128930	34.089	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051424\
Data File : VX041514.D
Acq On : 14 May 2024 12:53
Operator : JC/MD
Sample : PB160869TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB160869TB

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/15/2024
Supervised By :Mahesh Dadoda 05/15/2024

Quant Time: May 15 00:58:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

