

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051424\
 Data File : VX041515.D
 Acq On : 14 May 2024 13:16
 Operator : JC/MD
 Sample : P2444-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-DD

Quant Time: May 15 00:59:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	173776	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	259517	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	249068	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	123439	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	99281	52.480	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.960%
35) Dibromofluoromethane	5.379	113	80947	46.027	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.060%
50) Toluene-d8	8.647	98	304101	54.632	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.260%
62) 4-Bromofluorobenzene	11.079	95	113369	52.613	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.220%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	47305	53.712	ug/l	97
18) Methyl Acetate	2.715	43	3768	1.947	ug/l	96
20) Methylene Chloride	2.788	84	13894	8.460	ug/l	94
43) Isopropyl Acetate	6.342	43	118051	29.923	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051424\
Data File : VX041515.D
Acq On : 14 May 2024 13:16
Operator : JC/MD
Sample : P2444-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-DD

Quant Time: May 15 00:59:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

