

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051424\
 Data File : VX041518.D
 Acq On : 14 May 2024 14:25
 Operator : JC/MD
 Sample : P2470-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WCS-TPG

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/15/2024
 Supervised By : Mahesh Dadoda 05/15/2024

Quant Time: May 15 01:00:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 08 06:03:16 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	162742	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	244777	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240451	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112394	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92726	52.338	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 104.680%			
35) Dibromofluoromethane	5.385	113	76300	45.998	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 92.000%			
50) Toluene-d8	8.646	98	290011	55.238	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 110.480%			
62) 4-Bromofluorobenzene	11.079	95	105927	52.120	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 104.240%			
Target Compounds						
16) Acetone	2.385	43	40714	49.041	ug/l	97
18) Methyl Acetate	2.708	43	5545	3.059	ug/l	98
20) Methylene Chloride	2.788	84	11292	7.342	ug/l	93
37) Ethyl Acetate	4.745	43	3426m	1.384	ug/l	
43) Isopropyl Acetate	6.342	43	121261	32.587	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051424\
Data File : VX041518.D
Acq On : 14 May 2024 14:25
Operator : JC/MD
Sample : P2470-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WCS-TPG

Quant Time: May 15 01:00:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 08 06:03:16 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/15/2024
Supervised By : Mahesh Dadoda 05/15/2024

