

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051425\
 Data File : VX046183.D
 Acq On : 14 May 2025 10:50
 Operator : JC/MD
 Sample : VX0514WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0514WBL01

Quant Time: May 15 01:18:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	66143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	133003	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	124085	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54666	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	66199	53.684	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.360%	
35) Dibromofluoromethane	5.379	113	49039	51.202	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.400%	
50) Toluene-d8	8.647	98	165941	50.058	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.120%	
62) 4-Bromofluorobenzene	11.079	95	64922	51.057	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 102.120%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051425\
Data File : VX046183.D
Acq On : 14 May 2025 10:50
Operator : JC/MD
Sample : VX0514WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0514WBL01

Quant Time: May 15 01:18:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

