

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051425\
 Data File : VX046190.D
 Acq On : 14 May 2025 13:35
 Operator : JC/MD
 Sample : Q2001-05DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-A1-05-GDL

Quant Time: May 15 01:22:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	65402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130804	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	123296	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	57275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64073	52.549	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.100%
35) Dibromofluoromethane	5.379	113	47116	50.021	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.040%
50) Toluene-d8	8.647	98	164731	50.529	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.060%
62) 4-Bromofluorobenzene	11.079	95	67086	53.646	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.300%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	1852	2.050	ug/l #	87
16) Acetone	2.380	43	6054	12.383	ug/l	97
27) cis-1,2-Dichloroethene	4.483	96	22054	23.490	ug/l	88
40) Benzene	6.031	78	138974	37.490	ug/l	100
52) Toluene	8.714	92	87011	38.279	ug/l	98
67) Ethyl Benzene	10.189	91	279445	58.744	ug/l	100
68) m/p-Xylenes	10.299	106	36975	21.252	ug/l	95
69) o-Xylene	10.640	106	26489	15.617	ug/l	94
73) Isopropylbenzene	10.963	105	10433	2.340	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	8345	2.240	ug/l	96
84) 1,2,4-Trimethylbenzene	11.750	105	28813	7.638	ug/l	98
95) Naphthalene	13.774	128	1972992	494.009	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051425\
Data File : VX046190.D
Acq On : 14 May 2025 13:35
Operator : JC/MD
Sample : Q2001-05DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-A1-05-GDL

Quant Time: May 15 01:22:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

