

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051425\
 Data File : VX046193.D
 Acq On : 14 May 2025 14:45
 Operator : JC/MD
 Sample : Q2018-04DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW2DL

Quant Time: May 15 01:23:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	71196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	139917	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59001	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68079	51.291	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.580%		
35) Dibromofluoromethane	5.379	113	50731	50.351	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.700%		
50) Toluene-d8	8.647	98	176120	50.504	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.000%		
62) 4-Bromofluorobenzene	11.079	95	68773	51.413	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 102.820%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.965	59	34055	182.781	ug/l	98
16) Acetone	2.386	43	449	0.844	ug/l	97
39) Methylcyclohexane	7.379	83	480	0.275	ug/l #	81
40) Benzene	6.037	78	5105	1.287	ug/l #	89
73) Isopropylbenzene	10.963	105	6043	1.316	ug/l	97
78) n-propylbenzene	11.305	91	15400	2.883	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051425\
Data File : VX046193.D
Acq On : 14 May 2025 14:45
Operator : JC/MD
Sample : Q2018-04DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW2DL

Quant Time: May 15 01:23:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

