

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051425\
 Data File : VX046194.D
 Acq On : 14 May 2025 15:09
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: May 15 01:23:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	73494	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	145093	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	137314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60770	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	71408	52.116	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.240%	
35) Dibromofluoromethane	5.379	113	52741	50.478	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.960%	
50) Toluene-d8	8.647	98	181590	50.215	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.420%	
62) 4-Bromofluorobenzene	11.079	95	71333	51.424	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 102.840%	
Target Compounds						
					Qvalue	
95) Naphthalene	13.786	128	4468	1.054	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051425\
Data File : VX046194.D
Acq On : 14 May 2025 15:09
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: May 15 01:23:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

