

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051519\
 Data File : VX009576.D
 Acq On : 15 May 2019 13:40
 Operator : JC/SP
 Sample : K2838-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FEQC-FB-051419

Quant Time: May 16 01:42:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	188335	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	297510	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	262642	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	115334	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	123202	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
35) Dibromofluoromethane	5.49	113	91009	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
50) Toluene-d8	8.71	98	376494	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
62) 4-Bromofluorobenzene	11.13	95	125220	45.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.28%	
Target Compounds						
16) Acetone	2.44	43	8441	6.114	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051519\
Data File : VX009576.D
Acq On : 15 May 2019 13:40
Operator : JC/SP
Sample : K2838-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FEQC-FB-051419

Quant Time: May 16 01:42:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

