

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051519\
 Data File : VX009582.D
 Acq On : 15 May 2019 16:01
 Operator : JC/SP
 Sample : K2856-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS036-190514

Quant Time: May 16 02:01:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	181840	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	291500	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	256509	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	110645	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119904	48.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.60%	
35) Dibromofluoromethane	5.50	113	89120	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
50) Toluene-d8	8.71	98	364249	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
62) 4-Bromofluorobenzene	11.13	95	120604	44.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.72%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	34162	15.468	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051519\
Data File : VX009582.D
Acq On : 15 May 2019 16:01
Operator : JC/SP
Sample : K2856-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS036-190514

Quant Time: May 16 02:01:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

