

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051519\
 Data File : VX009583.D
 Acq On : 15 May 2019 16:24
 Operator : JC/SP
 Sample : K2857-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6B_R2

Quant Time: May 16 02:07:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	182101	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	294349	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	259731	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	110768	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	121153	48.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.48%	
35) Dibromofluoromethane	5.50	113	89456	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
50) Toluene-d8	8.71	98	369077	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
62) 4-Bromofluorobenzene	11.13	95	120697	44.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.92%	
Target Compounds						
16) Acetone	2.44	43	3236	2.424	ug/l #	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051519\
Data File : VX009583.D
Acq On : 15 May 2019 16:24
Operator : JC/SP
Sample : K2857-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6B_R2

Quant Time: May 16 02:07:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

