

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051519\
 Data File : VX009584.D
 Acq On : 15 May 2019 16:47
 Operator : JC/SP
 Sample : K2801-01
 Misc : 5.62g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RT-1873

Quant Time: May 15 17:08:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	157983	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.85	114	247539	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	212786	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	98178	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	102736	47.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.28%	
35) Dibromofluoromethane	5.50	113	70639	45.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.78%	
50) Toluene-d8	8.71	98	307192	48.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.80%	
62) 4-Bromofluorobenzene	11.14	95	103881	45.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.02%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051519\
Data File : VX009584.D
Acq On : 15 May 2019 16:47
Operator : JC/SP
Sample : K2801-01
Misc : 5.62µ/5mL/100µL/5.00mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RT-1873

Quant Time: May 15 17:08:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

