

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051519\
 Data File : VX009586.D
 Acq On : 15 May 2019 17:34
 Operator : JC/SP
 Sample : K2857-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-8A_R2

Quant Time: May 16 02:13:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	179430	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	286542	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	251151	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	108425	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	118379	48.33	ug/l	0.00
Spiked Amount			Recovery	=	96.66%	
35) Dibromofluoromethane	5.50	113	85361	47.39	ug/l	0.00
Spiked Amount			Recovery	=	94.78%	
50) Toluene-d8	8.71	98	355941	48.94	ug/l	0.00
Spiked Amount			Recovery	=	97.88%	
62) 4-Bromofluorobenzene	11.13	95	118528	44.85	ug/l	0.00
Spiked Amount			Recovery	=	89.70%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051519\
Data File : VX009586.D
Acq On : 15 May 2019 17:34
Operator : JC/SP
Sample : K2857-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8A_R2

Quant Time: May 16 02:13:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

