

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051520\
 Data File : VX016277.D
 Acq On : 15 May 2020 13:54
 Operator : JC/SP
 Sample : L2642-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 400513792-TRIP-BLANK

Quant Time: May 16 07:40:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	258198	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	423124	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	420249	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	211459	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	154012	45.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.66%	
35) Dibromofluoromethane	5.47	113	123656	46.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.26%	
50) Toluene-d8	8.70	98	530530	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
62) 4-Bromofluorobenzene	11.12	95	211645	53.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.74%	
Target Compounds						
16) Acetone	2.42	43	104826	69.147	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051520\
Data File : VX016277.D
Acq On : 15 May 2020 13:54
Operator : JC/SP
Sample : L2642-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
400513792-TRIP-BLANK

Quant Time: May 16 07:40:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

