

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051520\
 Data File : VX016278.D
 Acq On : 15 May 2020 14:17
 Operator : JC/SP
 Sample : L2665-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 0373-FIELD-BLANK

Quant Time: May 16 07:45:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	242599	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	403247	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	396908	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	199803	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	148522	46.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.04%	
35) Dibromofluoromethane	5.47	113	117489	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
50) Toluene-d8	8.70	98	501202	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
62) 4-Bromofluorobenzene	11.12	95	197672	52.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.60%	
Target Compounds						
16) Acetone	2.42	43	10785	7.572	ug/l	100
25) 2-Butanone	4.65	43	3348	1.459	ug/l	99
29) Tetrahydrofuran	5.09	42	6848	4.503	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051520\
Data File : VX016278.D
Acq On : 15 May 2020 14:17
Operator : JC/SP
Sample : L2665-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
0373-FIELD-BLANK

Quant Time: May 16 07:45:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

