

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051520\
 Data File : VX016284.D
 Acq On : 15 May 2020 16:34
 Operator : JC/SP
 Sample : L2649-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0684

Quant Time: May 16 07:58:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	239645	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	398286	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	395532	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	203169	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	145133	46.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.04%	
35) Dibromofluoromethane	5.47	113	116531	46.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.36%	
50) Toluene-d8	8.70	98	490552	50.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.92%	
62) 4-Bromofluorobenzene	11.12	95	198843	53.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.54%	
Target Compounds						
2) Dichlorodifluoromethane	1.18	85	5998	2.365	ug/l	100
4) Vinyl Chloride	1.39	62	4867	1.549	ug/l	99
16) Acetone	2.42	43	3114	2.213	ug/l	96
27) cis-1,2-Dichloroethene	4.57	96	2935	0.967	ug/l	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051520\
Data File : VX016284.D
Acq On : 15 May 2020 16:34
Operator : JC/SP
Sample : L2649-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0684

Quant Time: May 16 07:58:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

