

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051523\
 Data File : VX035660.D
 Acq On : 15 May 2023 13:16
 Operator : JC/MD
 Sample : 02751-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-PRO-01-89-05062023

Quant Time: May 16 04:58:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	188535	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	341750	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	308532	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138997	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	154961	47.114	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.220%
35) Dibromofluoromethane	5.385	113	112281	48.717	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.440%
50) Toluene-d8	8.647	98	408887	48.023	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.040%
62) 4-Bromofluorobenzene	11.079	95	167879	48.970	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.940%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.318	96	2307	1.099	ug/l	93
16) Acetone	2.398	43	7609	5.383	ug/l	98
17) Carbon Disulfide	2.514	76	1752	0.346	ug/l #	91
21) trans-1,2-Dichloroethene	3.093	96	1484	0.645	ug/l	82
27) cis-1,2-Dichloroethene	4.483	96	3977	1.438	ug/l	94
44) Trichloroethene	7.129	130	1774984	687.830	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051523\
Data File : VX035660.D
Acq On : 15 May 2023 13:16
Operator : JC/MD
Sample : 02751-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-PRO-01-89-05062023

Quant Time: May 16 04:58:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

