

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051523\
 Data File : VX035668.D
 Acq On : 15 May 2023 16:22
 Operator : JC/MD
 Sample : 02751-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-DUP-02-05062023

Quant Time: May 16 05:01:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	189251	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	349747	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	316027	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	143862	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	157944	47.839	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.680%
35) Dibromofluoromethane	5.385	113	111887	47.436	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.880%
50) Toluene-d8	8.647	98	419257	48.115	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.240%
62) 4-Bromofluorobenzene	11.079	95	171377	48.847	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.700%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.325	96	2585	1.227	ug/l	82
16) Acetone	2.398	43	4840	3.411	ug/l	98
44) Trichloroethene	7.129	130	2598	0.984	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 16 05:01:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

