

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051523\
 Data File : VX035669.D
 Acq On : 15 May 2023 16:45
 Operator : JC/MD
 Sample : 02751-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-EQB-03-05052023

Quant Time: May 16 05:01:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	184200	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	341785	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	304899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137750	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	154059	47.942	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.880%
35) Dibromofluoromethane	5.385	113	109699	47.592	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.180%
50) Toluene-d8	8.647	98	408895	48.019	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.040%
62) 4-Bromofluorobenzene	11.079	95	166480	48.557	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.120%
Target Compounds						
					Qvalue	
16) Acetone	2.404	43	3316	2.401	ug/l	98
29) Tetrahydrofuran	5.038	42	2588	2.154	ug/l	# 76
37) Ethyl Acetate	4.715	43	41777	11.009	ug/l	99
44) Trichloroethene	7.129	130	1335	0.517	ug/l	96
52) Toluene	8.720	92	1596	0.248	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051523\
Data File : VX035669.D
Acq On : 15 May 2023 16:45
Operator : JC/MD
Sample : 02751-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-EQB-03-05052023

Quant Time: May 16 05:01:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

