

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051523\
 Data File : VX035670.D
 Acq On : 15 May 2023 17:09
 Operator : JC/MD
 Sample : 02751-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-EQB-04-05062023

Quant Time: May 16 05:02:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	199927	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	369544	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	341959	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	156634	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	166257	47.668	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.340%
35) Dibromofluoromethane	5.385	113	117299	47.067	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.140%
50) Toluene-d8	8.647	98	443508	48.172	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.340%
62) 4-Bromofluorobenzene	11.079	95	185658	50.082	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.160%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	3146	2.099	ug/l	# 87
29) Tetrahydrofuran	5.032	42	2420	1.855	ug/l	97
37) Ethyl Acetate	4.721	43	38578	9.402	ug/l	99
44) Trichloroethene	7.123	130	1281	0.459	ug/l	92
52) Toluene	8.720	92	1590	0.229	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051523\
Data File : VX035670.D
Acq On : 15 May 2023 17:09
Operator : JC/MD
Sample : 02751-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-EQB-04-05062023

Quant Time: May 16 05:02:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

