

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051523\
 Data File : VX035672.D
 Acq On : 15 May 2023 17:55
 Operator : JC/MD
 Sample : 02751-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-EQB-06-05082023

Quant Time: May 16 05:02:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 11 13:57:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	186220	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	341497	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	315405	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140172	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156676	48.228	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.460%		
35) Dibromofluoromethane	5.385	113	111685	48.494	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.980%		
50) Toluene-d8	8.646	98	411875	48.410	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.820%		
62) 4-Bromofluorobenzene	11.079	95	167222	48.814	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.620%		
Target Compounds					Qvalue	
16) Acetone	2.404	43	4279	3.065	ug/l	93
37) Ethyl Acetate	4.714	43	42903	11.315	ug/l	99
44) Trichloroethene	7.122	130	684	0.265	ug/l	52

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051523\
Data File : VX035672.D
Acq On : 15 May 2023 17:55
Operator : JC/MD
Sample : 02751-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-EQB-06-05082023

Quant Time: May 16 05:02:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051123W.M
Quant Title : SW846 8260
QLast Update : Thu May 11 13:57:54 2023
Response via : Initial Calibration

