



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051524\
Data File : VX041527.D
Acq On : 15 May 2024 08:32
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Thu May 16 05:13:13 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	10079	PASS
75	95	30	60	50.1	25808	PASS
95	95	100	100	100.0	51533	PASS
96	95	5	9	7.1	3660	PASS
173	174	0.00	2	1.3	615	PASS
174	95	50	100	94.4	48629	PASS
175	174	5	9	7.6	3705	PASS
176	174	95	101	96.5	46941	PASS
177	176	5	9	6.6	3119	PASS