

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
 Data File : VX046208.D
 Acq On : 15 May 2025 13:01
 Operator : JC/MD
 Sample : Q2050-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-60-62

Quant Time: May 16 01:04:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	61866	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	123143	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50521	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63133	54.737	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.480%		
35) Dibromofluoromethane	5.379	113	45470	51.277	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.560%		
50) Toluene-d8	8.647	98	152829	49.795	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.580%		
62) 4-Bromofluorobenzene	11.079	95	58894	50.025	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 100.040%		
Target Compounds						
						Qvalue
16) Acetone	2.380	43	3602	7.789	ug/l	96
30) Chloroform	5.099	83	1681	1.069	ug/l	97
64) Tetrachloroethene	9.281	164	1022	1.255	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
Data File : VX046208.D
Acq On : 15 May 2025 13:01
Operator : JC/MD
Sample : Q2050-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-60-62

Quant Time: May 16 01:04:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

