

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
 Data File : VX046211.D
 Acq On : 15 May 2025 14:11
 Operator : JC/MD
 Sample : Q2050-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-100-102

Quant Time: May 16 01:05:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	61221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	120798	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115485	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	47647	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	62643	54.885	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.760%		
35) Dibromofluoromethane	5.379	113	46065	52.956	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.920%		
50) Toluene-d8	8.647	98	154318	51.256	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.520%		
62) 4-Bromofluorobenzene	11.079	95	57300	49.616	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 99.240%		
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1754	3.833	ug/l	93
30) Chloroform	5.093	83	1439	0.925	ug/l	97
64) Tetrachloroethene	9.275	164	928	1.136	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
Data File : VX046211.D
Acq On : 15 May 2025 14:11
Operator : JC/MD
Sample : Q2050-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-100-102

Quant Time: May 16 01:05:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

