

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
 Data File : VX046212.D
 Acq On : 15 May 2025 14:34
 Operator : JC/MD
 Sample : Q2050-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-150-152

Quant Time: May 16 01:06:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	63694	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128038	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	119578	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51539	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66083	55.651	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.300%
35) Dibromofluoromethane	5.385	113	48275	52.358	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.720%
50) Toluene-d8	8.647	98	161234	50.525	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.040%
62) 4-Bromofluorobenzene	11.079	95	61266	50.050	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.100%
Target Compounds						
16) Acetone	2.380	43	1902	3.995	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
Data File : VX046212.D
Acq On : 15 May 2025 14:34
Operator : JC/MD
Sample : Q2050-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-150-152

Quant Time: May 16 01:06:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

