

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
 Data File : VX046213.D
 Acq On : 15 May 2025 14:58
 Operator : JC/MD
 Sample : Q2050-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-205-207

Quant Time: May 16 01:06:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	60533	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	121144	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	47407	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62575	55.448	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 110.900%			
35) Dibromofluoromethane	5.379	113	45370	52.008	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.020%			
50) Toluene-d8	8.647	98	155555	51.519	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.040%			
62) 4-Bromofluorobenzene	11.079	95	57300	49.474	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 98.940%			
Target Compounds						
16) Acetone	2.380	43	2332	5.154	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
Data File : VX046213.D
Acq On : 15 May 2025 14:58
Operator : JC/MD
Sample : Q2050-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-205-207

Quant Time: May 16 01:06:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

