

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
 Data File : VX046215.D
 Acq On : 15 May 2025 15:44
 Operator : JC/MD
 Sample : Q2050-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-DUP-20250514

Quant Time: May 16 01:07:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	61073	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	122275	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	116802	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50006	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63331	55.622	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.240%
35) Dibromofluoromethane	5.379	113	45823	52.041	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.080%
50) Toluene-d8	8.647	98	154226	50.606	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.220%
62) 4-Bromofluorobenzene	11.079	95	60054	51.372	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.740%
Target Compounds						
16) Acetone	2.380	43	1288	2.821	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
Data File : VX046215.D
Acq On : 15 May 2025 15:44
Operator : JC/MD
Sample : Q2050-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-DUP-20250514

Quant Time: May 16 01:07:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

