

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
 Data File : VX046218.D
 Acq On : 15 May 2025 16:54
 Operator : JC/MD
 Sample : Q2034-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 L3-WC-4

Quant Time: May 16 01:08:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	62053	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	122291	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	119697	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52624	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63161	54.597	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.200%	
35) Dibromofluoromethane	5.379	113	45614	51.797	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.600%	
50) Toluene-d8	8.647	98	159242	52.245	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.500%	
62) 4-Bromofluorobenzene	11.079	95	62227	53.224	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 106.440%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	17549	37.833	ug/l	96
18) Methyl Acetate	2.715	43	2231	2.073	ug/l	95
20) Methylene Chloride	2.788	84	14627	16.455	ug/l	99
25) 2-Butanone	4.586	43	5169	7.676	ug/l #	81
43) Isopropyl Acetate	6.348	43	12548	5.627	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
Data File : VX046218.D
Acq On : 15 May 2025 16:54
Operator : JC/MD
Sample : Q2034-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
L3-WC-4

Quant Time: May 16 01:08:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

