

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
 Data File : VX046220.D
 Acq On : 15 May 2025 17:41
 Operator : JC/MD
 Sample : Q2032-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-1

Quant Time: May 16 01:09:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	60792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124950	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117007	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49771	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63320	55.869	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.740%
35) Dibromofluoromethane	5.379	113	45972	51.093	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.180%
50) Toluene-d8	8.647	98	158281	50.825	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.660%
62) 4-Bromofluorobenzene	11.079	95	58875	49.285	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.580%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	25113	55.263	ug/l	100
18) Methyl Acetate	2.703	43	2985	2.831	ug/l	95
20) Methylene Chloride	2.782	84	17853	20.501	ug/l	96
25) 2-Butanone	4.574	43	6323	9.584	ug/l	93
43) Isopropyl Acetate	6.342	43	13487	5.919	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
Data File : VX046220.D
Acq On : 15 May 2025 17:41
Operator : JC/MD
Sample : Q2032-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-1

Quant Time: May 16 01:09:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

