

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
 Data File : VX046221.D
 Acq On : 15 May 2025 18:04
 Operator : JC/MD
 Sample : Q2034-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 L3-WC-1

Quant Time: May 16 01:09:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	62263	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126954	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	123968	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51188	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64665	55.708	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.420%
35) Dibromofluoromethane	5.385	113	46875	51.274	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.540%
50) Toluene-d8	8.647	98	163601	51.704	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.400%
62) 4-Bromofluorobenzene	11.079	95	63189	52.062	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.120%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	19161	41.169	ug/l	96
18) Methyl Acetate	2.709	43	3590	3.324	ug/l	97
20) Methylene Chloride	2.782	84	14689	16.469	ug/l	98
25) 2-Butanone	4.586	43	5549	8.212	ug/l	99
43) Isopropyl Acetate	6.342	43	13415	5.794	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
Data File : VX046221.D
Acq On : 15 May 2025 18:04
Operator : JC/MD
Sample : Q2034-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
L3-WC-1

Quant Time: May 16 01:09:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

