

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
 Data File : VX046223.D
 Acq On : 15 May 2025 18:51
 Operator : JC/MD
 Sample : Q2034-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 L3-WC-3

Quant Time: May 16 01:10:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	61870	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	125392	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	122779	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52539	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63329	54.904	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.800%	
35) Dibromofluoromethane	5.379	113	46298	51.274	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.540%	
50) Toluene-d8	8.647	98	160472	51.347	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.700%	
62) 4-Bromofluorobenzene	11.079	95	62276	51.949	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.900%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	17893	38.689	ug/l	99
18) Methyl Acetate	2.709	43	3486	3.248	ug/l	97
20) Methylene Chloride	2.782	84	12690	14.319	ug/l	99
25) 2-Butanone	4.581	43	6233	9.283	ug/l	96
43) Isopropyl Acetate	6.342	43	14259	6.236	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051525\
Data File : VX046223.D
Acq On : 15 May 2025 18:51
Operator : JC/MD
Sample : Q2034-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
L3-WC-3

Quant Time: May 16 01:10:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

