

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051619\
 Data File : VX009609.D
 Acq On : 16 May 2019 15:57
 Operator : JC/SP
 Sample : K2886-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-WK-3

Quant Time: May 17 07:04:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	176035	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	275295	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	235269	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	93169	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	114586	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
35) Dibromofluoromethane	5.50	113	84508	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
50) Toluene-d8	8.71	98	346447	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
62) 4-Bromofluorobenzene	11.13	95	107445	42.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.64%	
Target Compounds						
16) Acetone	2.45	43	3926	3.043	ug/l	89
18) Methyl Acetate	2.78	43	1791	0.584	ug/l #	88
27) cis-1,2-Dichloroethene	4.59	96	28191	13.185	ug/l	90
29) Tetrahydrofuran	5.15	42	6807	5.275	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051619\
Data File : VX009609.D
Acq On : 16 May 2019 15:57
Operator : JC/SP
Sample : K2886-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-WK-3

Quant Time: May 17 07:04:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

