

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051619\
 Data File : VX009626.D
 Acq On : 16 May 2019 23:23
 Operator : JC/SP
 Sample : K2887-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z4-TB06-190514

Quant Time: May 17 07:39:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	169410	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	268270	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	233351	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	95653	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	112313	48.56	ug/l	0.00
Spiked Amount			Recovery	=	97.12%	
35) Dibromofluoromethane	5.50	113	82954	49.19	ug/l	0.00
Spiked Amount			Recovery	=	98.38%	
50) Toluene-d8	8.71	98	335556	49.28	ug/l	0.00
Spiked Amount			Recovery	=	98.56%	
62) 4-Bromofluorobenzene	11.13	95	109553	44.28	ug/l	0.00
Spiked Amount			Recovery	=	88.56%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051619\
Data File : VX009626.D
Acq On : 16 May 2019 23:23
Operator : JC/SP
Sample : K2887-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z4-TB06-190514

Quant Time: May 17 07:39:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

