

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051619\
 Data File : VX009627.D
 Acq On : 16 May 2019 23:46
 Operator : JC/SP
 Sample : K2887-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z4-TB07-190514

Quant Time: May 17 07:40:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	169401	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	269099	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	234018	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	98645	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	112935	48.84	ug/l	0.00
Spiked Amount			Recovery	=	97.68%	
35) Dibromofluoromethane	5.50	113	82413	48.72	ug/l	0.00
Spiked Amount			Recovery	=	97.44%	
50) Toluene-d8	8.71	98	334069	48.91	ug/l	0.00
Spiked Amount			Recovery	=	97.82%	
62) 4-Bromofluorobenzene	11.13	95	109639	44.18	ug/l	0.00
Spiked Amount			Recovery	=	88.36%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051619\
Data File : VX009627.D
Acq On : 16 May 2019 23:46
Operator : JC/SP
Sample : K2887-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z4-TB07-190514

Quant Time: May 17 07:40:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

