

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051619\
 Data File : VX009630.D
 Acq On : 17 May 2019 00:56
 Operator : JC/SP
 Sample : K2887-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW128-190514

Quant Time: May 17 07:46:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	173201	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	275779	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	244335	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	104302	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	114716	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
35) Dibromofluoromethane	5.50	113	83879	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
50) Toluene-d8	8.71	98	345304	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
62) 4-Bromofluorobenzene	11.13	95	114449	45.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.00%	
Target Compounds						
4) Vinyl Chloride	1.40	62	3127	1.472	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	22288	10.595	ug/l	88
44) Trichloroethene	7.21	130	2491	1.299	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051619\
Data File : VX009630.D
Acq On : 17 May 2019 00:56
Operator : JC/SP
Sample : K2887-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW128-190514

Quant Time: May 17 07:46:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

