

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051619\
 Data File : VX009634.D
 Acq On : 17 May 2019 02:30
 Operator : JC/SP
 Sample : K2887-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW733-190515

Quant Time: May 17 07:52:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	171103	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	272323	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	238802	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	99474	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	115089	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
35) Dibromofluoromethane	5.50	113	83590	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
50) Toluene-d8	8.71	98	341889	49.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.94%	
62) 4-Bromofluorobenzene	11.13	95	111583	44.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.86%	
Target Compounds						
12) 1,1-Dichloroethene	2.37	96	9704	6.044	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	3297	1.586	ug/l	94
44) Trichloroethene	7.21	130	2248	1.187	ug/l	100
64) Tetrachloroethene	9.34	164	13567	8.282	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051619\
Data File : VX009634.D
Acq On : 17 May 2019 02:30
Operator : JC/SP
Sample : K2887-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW733-190515

Quant Time: May 17 07:52:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

