

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX051619\
 Data File : VX009638.D
 Acq On : 17 May 2019 04:04
 Operator : JC/SP
 Sample : K2887-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS052MW098-190514

Quant Time: May 17 07:57:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 09 07:31:15 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	173172	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	280164	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	249467	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	106946	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117083	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
35) Dibromofluoromethane	5.50	113	85566	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
50) Toluene-d8	8.71	98	350570	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
62) 4-Bromofluorobenzene	11.13	95	117770	45.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.16%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	5320	2.529	ug/l	87
44) Trichloroethene	7.21	130	22722	11.664	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX051619\
Data File : VX009638.D
Acq On : 17 May 2019 04:04
Operator : JC/SP
Sample : K2887-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS052MW098-190514

Quant Time: May 17 07:57:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Thu May 09 07:31:15 2019
Response via : Initial Calibration

