

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
 Data File : VX028724.D
 Acq On : 16 May 2022 12:24
 Operator : JC/MD
 Sample : N2813-05 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BUR-1149

Quant Time: May 17 03:34:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	242345	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	443597	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	423707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	193118	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	172408	51.899	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.800%	
35) Dibromofluoromethane	5.391	113	149444	51.419	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.840%	
50) Toluene-d8	8.653	98	536188	48.933	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 97.860%	
62) 4-Bromofluorobenzene	11.085	95	199095	48.646	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.300%	
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	81106	56.888	ug/l	95
25) 2-Butanone	4.574	43	3862	1.726	ug/l	90
52) Toluene	8.726	92	4241	0.525	ug/l	100
68) m/p-Xylenes	10.299	106	1874	0.299	ug/l	92
84) 1,2,4-Trimethylbenzene	11.756	105	6150	0.503	ug/l	97
95) Naphthalene	13.780	128	13596	0.991	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
Data File : VX028724.D
Acq On : 16 May 2022 12:24
Operator : JC/MD
Sample : N2813-05 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BUR-1149

Quant Time: May 17 03:34:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

