

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
 Data File : VX028726.D
 Acq On : 16 May 2022 13:11
 Operator : JC/MD
 Sample : N2862-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB051222

Quant Time: May 17 03:34:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	237334	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	432871	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	420813	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	217050	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	169835	52.204	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.400%	
35) Dibromofluoromethane	5.385	113	144293	50.877	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.760%	
50) Toluene-d8	8.653	98	527661	49.348	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.700%	
62) 4-Bromofluorobenzene	11.085	95	196397	49.176	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.360%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.947	59	16080	15.816	ug/l #	73
16) Acetone	2.380	43	3975	2.847	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
Data File : VX028726.D
Acq On : 16 May 2022 13:11
Operator : JC/MD
Sample : N2862-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB051222

Quant Time: May 17 03:34:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

