

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
 Data File : VX028727.D
 Acq On : 16 May 2022 13:34
 Operator : JC/MD
 Sample : N2862-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB051322

Quant Time: May 17 03:35:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	238448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	439158	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	419917	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	185250	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	172084	52.648	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.300%
35) Dibromofluoromethane	5.391	113	143861	49.999	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.000%
50) Toluene-d8	8.653	98	532892	49.124	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.240%
62) 4-Bromofluorobenzene	11.079	95	194720	48.058	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.120%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.941	59	16836	16.830	ug/l #	76
16) Acetone	2.374	43	2536	1.808	ug/l	93
20) Methylene Chloride	2.788	84	1161	0.394	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
Data File : VX028727.D
Acq On : 16 May 2022 13:34
Operator : JC/MD
Sample : N2862-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB051322

Quant Time: May 17 03:35:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

