

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
 Data File : VX028728.D
 Acq On : 16 May 2022 13:58
 Operator : JC/MD
 Sample : N2862-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1R

Quant Time: May 17 03:35:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 12 14:58:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	226153	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	417423	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	393620	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	170300	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	164131	52.945	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.880%
35) Dibromofluoromethane	5.391	113	139835	51.130	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.260%
50) Toluene-d8	8.653	98	501034	48.592	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.180%
62) 4-Bromofluorobenzene	11.079	95	180226	46.797	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.600%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	2511	1.887	ug/l	91
30) Chloroform	5.098	83	5308	1.052	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051622\
Data File : VX028728.D
Acq On : 16 May 2022 13:58
Operator : JC/MD
Sample : N2862-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1R

Quant Time: May 17 03:35:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051222W.M
Quant Title : SW846 8260
QLast Update : Thu May 12 14:58:53 2022
Response via : Initial Calibration

